Developing a Logarithmic Chinese Reading Acuity Chart

Qi-Ming Han*, Lin-Juan Cong[†], Cong Yu[†], and Lei Liu[‡]

ABSTRACT

Purpose. A h /h ga ิล Čh Ch ha FAD) h ga ha h а 12- ha a Methods. h h Ch \mathbf{O} ิล ± 0.05 g C-FAD h Α а hınBh ha h C- EAD ag hſ ŕB а • **1** 34.37 Results. Ch g a 0.24 ± 0.06 273.44 ± haja 0.16 ± 0.05 a g h h C-EAD a ha 11 h nB h , ha a FAD h Ch a Conclusions. h D ha a a a a h a a a g а ิล a :714-724) (O)

∙,:C-EAD, arga ha, forch

dedicated instrument to assess a person's ability to read regular text is needed because this ability cannot be reliably predicted by letter acuity, contrast sensitivity, and visual field extent.^{1,2} There has been a long history of developing continuous text reading tests and applying them to vision care.^{3–7} The Minnesota Low-Vision Reading Test (MNRead) and the Radner Reading charts are the two most notable ones.^{5,6,8} Both tests characterize reading performance by reading acuity, critical print size, and maximum reading speed. They share some time-tested design principles, such as standardized continuous text test items that closely resemble everyday reading materials, high-frequency vocabulary at the third-grade reading level, most popular typefaces, logarithmic progression of print sizes, and uniformity in text spatial layout,^{5,8} but they also differ in several ways. The MNRead uses simple, 60-character (including spacing) declarative sentences shown in three lines.⁵ The German version of the Radner Reading charts uses "sentence optotypes," which are highly comparable in terms of the number of words per sentence and format (14 words, 82 to 84 characters, printed in three lines), the number of words per line (five words in lines 1 and 2, four words in line 3), the word length in specific sentence locations, the position of the relative clause (following the second word of line 2), and the distribution of syllables within a sentence.⁸

While these design principles have been adopted in developing reading charts in multiple languages,^{9–15} their applications to Chinese text may not be straightforward. The logographic Chinese differs from the linear alphabetic Latin languages in several ways, which on one hand demands modifications of some of the principles mentioned previously, but on the other hand makes implementation of other principles more natural. The following unique features of the Chinese text need to be considered in designing a Chinese reading test.

^{*}MPhil

[†]PhD

[‡]PhD, FAAO

Sentence Composition

The Chinese language lacks the relative pronouns, such as *who*, *which*, and *that*. Therefore, sentence compositions that are more complex than simple declarative sentences, such as those used in Radner Reading chart, may introduce uncertainties such as pauses between clauses and repetitions of subject nouns.

Sentence Length Measurement

The smallest meaningful unit of written Chinese is a character. It is therefore natural to measure sentence length in terms of characters instead of words. The IReST uses characters as the basis for measuring reading speed.^{16,17} All 60-character MNRead sentences are evaluated based on 10 "standard-length" (six characters) words,¹⁴ whereas each Radner Reading sentence contains 14 words.⁸

Physical Layout

Each sentence of the MNRead and Radner Reading charts is printed in three lines, 20 characters per line for MNRead and 24 or 25 characters per line for Radner Reading chart.^{8,14} Considering the printing practice of 65 characters or approximately 10 words per line for paperback English books, breaking these test sentences into three lines is not too far from normal reading practice.

Chinese text is more compact than Latin language texts.^{18,19} Chinese paper books have 28 to 32 characters per line. Breaking a 12-character sentence into three lines would not reflect normal Chinese reading habits. Using text lines this short also increases the risk of breaking a compound word and thus introduces unwanted pauses or hesitation.

All Chinese characters of the same typeface and font size occupy the same square area, and there is no spacing between words. Sentences that contain the same number of characters naturally have the same horizontal extent.

Number of Syllables

Each Chinese character is one syllable. Therefore, equalizing the number of characters in the testing sentences also equalizes the number of syllables, making quantification of verbal reading more accurate.

Inclusion of Simpler and More Complex Characters

The Latin alphabets are quite simple and relatively uniform in spatial complexity. In comparison, even Chinese text for beginners is a mixture of characters of very different spatial complexities. The number of strokes of the first 1000 most frequently used Chinese characters ranges from 1 (一, one) to 25 (露, dew). To reflect everyday reading performance, characters of all complexities should be included according to their natural occurrence. However, characters of different spatial complexities have very different legibility.²⁰ Moreover, consecutive simple characters tend to open a wide "gap" with less ink in a sentence that may alleviate crowding for the characters on the edges of the gap and thus facilitate character

recognition. Therefore, simple characters in test sentences need to be equalized and scattered in a similar manner.

Regional Dialects

There are many regional dialects in China that differ greatly. Some characters are more difficult to pronounce in some dialects than in others. There are also words and phrases that are more frequently used in some dialects. These differences may have direct impacts on verbal reading speed assessment.

Print Size Range

The MNRead covers an acuity range between 20/6.3 and 20/ 400. The Radner Reading chart covers an acuity range between 20/12.5 and 20/320. The single-character acuity sizes of Chinese characters are at least 0.1 log unit larger than the Sloan letter acuity size.²⁰ Therefore, a Chinese reading acuity chart may not need print sizes for 20/6.3 or 20/8.

This article describes the development and validation of a reading test for readers of simplified Chinese. This instrument, named the C-READ, is inspired by many of the design principles embodied in the MNRead and Radner Reading charts, but is also customized to accommodate the special needs of a Chinese reading test.

METHODS

The Development of a Simplified C-READ

At the beginning, 105 simple declarative sentences ("subject-verbobject") were crafted from the material of the first- to third-grade textbooks of Chinese elementary schools.²¹ Each sentence had 12 characters or syllables, which were comparable with the MNRead sentences (12 to 15 syllables, counted in www.wordcalc. com), but were significantly shorter than the German, Dutch, and Spanish versions of the Radner Reading charts (20 to 30 syllables).^{8,12,22} Reading a 12-character sentence at the maximum speed took 2.68 \pm 0.36 seconds, similar to that for reading a MNRead sentence (~2.97 seconds per sentence). Therefore, the chart would have similar testing time and incur a similar observer burden as the MNRead.

Next, sentences with too many or too few total number of strokes were eliminated, reducing the pool of sentences to 67. These sentences were tested for reading speed, reading error, and reading fluency in 20 Chinese college students. These students also subjectively rated the suitability of these sentences for reading assessment on a 1-point (least) to 5-point (most) scale. Sentences that contained characters that were easily confused with others in sounds, expressions that were age specific, or expressions that were not commonly used received lower scores. Sentences with the longest and shortest reading times and/or with the lowest suitability scores were removed until 48 sentences remained.

Finer adjustments to the sentences were then made to replace less frequently used characters and to limit the number of very simple characters (three strokes or less) in the sentences, so that each sentence had no more than three very simple characters and no three consecutive very simple characters. The mean number of strokes of the 283 unique characters included in these sentences was 7.13 (range, 1 to 15 strokes), which was fewer than the mean number of strokes of the 2570 characters listed in the textbooks of all six primary school grades (9.45 strokes).²³ This was because the sentences used for developing the C-READ were selected from the textbooks of grades 1 to 3 and because textbooks for higher grades contained more complex characters.

The final 48 sentences were randomly assigned to three 16sentence charts. Each chart covered a print size range from -0.3 to 1.2 logMAR (20/10 to 20/320) in 0.1 logMAR steps for a 40-cm reading distance. The mean strokes per sentence were 85.3 ± 3.5 , 85.6 ± 3.0 , and 85.9 ± 3.2 for charts A, B, and C, respectively. The character height was determined by the vertical extent of characters that have horizontal strokes on the top and bottom and have a sufficient number of strokes, such as \mathbb{A} and \mathbb{H} . The vector font size in points for each line of the chart was first adjusted to match the nominal character height (e.g., the nominal height of the 20/20 line characters is 5 arc minutes or 0.582 mm at 40 cm) in Adobe Illustrator during chart character rization. After the chart was printed on paper, the actual character heights were verified under a 15 times measuring loupe.

The reading charts used typeface Song Ti (



FIGURE 1.

Cha Alh ga h	C- EAD. h	ha a I	12-haja	.∎aa	,	ga a	a gl	20/10	20/320.
ha fah	a 📭 📲	aa a∙,	g A h f	a	fa f	40 a.∎	20	h gh ∎	

errors. The reading time was adjusted by the number of uncorrected errors before being converted to reading speed in characters per minute, as recommended by the developers of IReST.¹⁶ Each observer was also tested with two randomly selected versions of the C-READ with corrected and uncorrected near vision.

Data Analysis

Three algorithms for extracting C-READ reading performance parameters were compared. Both the original MNRead scoring algorithm and the two-limb algorithm used a sloped line to fit the increasing reading speed at the smaller print sizes and a horizontal line to fit the reading speed plateau at larger print sizes.^{14,24} The two line segments were determined separately in the MNRead algorithm but were fitted together based on one model in the two-limb algorithm. The third algorithm used an exponential-decay function to fit the entire reading speed curve.²⁴ The original MNRead algorithm defined reading acuity as the size of the last sentence attempted, plus the number of uncorrected word errors in that sentence timed 0.01 (10 words in each 0.1 logMAR line). The other two algorithms defined reading acuity as the intersection of the best-fitting line/curve with the horizontal axis (the print size when the reading speed dropped to zero).

These algorithms were modified to better fit the C-READ data. First, the MNRead assumed that each sentence contained 10 standard-length words (six characters per word), and each word was worth 0.01 log unit. Because each C-READ sentence had 12 characters, each character was worth 0.1/12 = 0.00833 log unit. Second, a "floor effect" was frequently observed in C-READ tests, in that the observers could only correctly read one or two characters in the two or three smallest sentences before they finally gave up.

differences for each of the three C-READ parameters.²⁵ The measurement errors for using the three charts as a set were quantified by within-subject SDs of the C-READ parameters.²⁶

A variance component analysis with individual observers and C-READ versions as the random factors, test order as the fixed variable, and reading acuity, critical print size, and maximum reading speed as dependent variables was conducted to determine the sources of variability in the C-READ parameters. Pearson *r* was calculated to assess the agreement between the C-READ and the IReST reading speed, as well as the relationship among C-READ parameters and uncorrected near acuity. A linear regression analysis with the IReST reading speed as the dependent variable and the C-READ parameters as the independent variables was performed to establish the relationship between the C-READ parameters and the continuous text reading performance.

The data are presented as mean \pm SD throughout the article.

RESULTS

Normative Data of the C-READ

Table 1 shows the reading parameters of reading acuity, critical print size, and maximum reading speed of individual charts and their averages obtained from 30 native Chinese-speaking colleague students. There were no significant differences in the reading parameters among the three charts ($F_{2,27} = 2.64$, P = .09, for reading acuity; F_{2,27} = 2.68, P = .09, for critical print size; and F_{2,27} = 1.93, P = .17, for maximum reading speed). Intraclass correlations among the reading parameters derived from the three versions of the C-READ were excellent (intraclass correlation = 0.98, 0.93, and 0.91 for reading acuity, critical print size, and maximum reading speed, respectively), suggesting good interchart reliability.²⁷ There was no significant test order effect of the same charts on the three parameters either ($F_{2,27} = 1.13$, P =.34, for reading acuity; $F_{2,27} = 1.21$, P = .32, for critical print size; and $F_{2,27} =$ 1.46, P = .25, for maximum reading speed). Intraclass correlations for testing order were good to excellent (intraclass correlation = 0.85, 0.72, and 0.99 for reading acuity, critical print size, and maximum reading speed, respectively), suggesting adequate test-retest reliability.

The pairwise Bland-Altman plots for the three reading parameters are shown in Fig. 3. For reading acuity, the 95% limits of agreement, the interval demarked by the pair of black dashed lines, were -0.004 ± 0.066 , -0.001 ± 0.047 , and 0.003 ± 0.061 for chart pairs A versus B, A versus C, and B versus C, respectively. If any pair of the C-READs is used for repeated tests, 95% of the measurement differences in reading acuity would be smaller than one line (0.1 logMAR). The assessments of the limits of agreement were accurate. The SEs (the dotted lines around each dashed line) were 0.011, 0.008, and 0.010 logMAR, respectively. Similarly, critical print sizes measured using any two versions of C-READ would not differ more

than two lines (the 95% limits of agreement were -0.011 ± 0.170 , -0.006 ± 0.162 , and 0.005 ± 0.167). The maximum reading speeds obtained with any two charts would not differ more than 27 characters per minute (95% limits of agreement were 0.709 ± 26.485 , -5.245 ± 24.612 , and -5.954 ± 26.775). If all three charts are used on the same subject, the within-subject SDs for reading acuity, critical print size, and maximum reading speed were 0.02 logMAR, 0.06 logMAR, and 9.8 characters per minute, respectively, suggesting good interchart reliability.²⁶

The variance component analysis revealed that the observers were the predominant factor influencing the C-READ variability, contributing 96.5%, 65.1%, and 81.7% to the total variances of reading acuity, critical print size, and maximum reading speed, respectively (Table 2). The test orders influenced maximum reading speed only, whereas the three versions of the C-READ did not influence any of the parameters. A considerable proportion of the critical print size and maximum reading speed variability came from the interaction between the observers and the charts (32.5% for critical print size, 10.1% for maximum reading speed). The interaction of charts and test orders also contributed a small proportion of variance to the three parameters. These data suggested that most of the chart parameter variances were caused by interobserver variability, not by the charts or the test orders.

The Relationship between C-READ Parameters and IReST Reading Speed

Fig. 4 shows that the observers in this experiment had uncorrected near acuities spreading over a 1-log-unit range. Table 3 shows that the IReST reading speed was significantly faster than C-READ maximum reading speed under both corrected and uncorrected near-vision conditions. These differences were highly significant ($t_{31} = 8.37$, P < .001, with corrected vision; $t_{31} = 3.22$, P = .003, with uncorrected vision; two-tailed paired *t* test). The IReST reading speed and the chart maximum reading speed were highly correlated when near vision was corrected (r = 0.87, P < .001) and were moderately correlated when near vision was not corrected (r = 0.59, P < .001).

Assessing IReST Reading Performance with C-READ at Various Refractive Errors

Fig. 5 is the scatter plot of the chart maximum reading speed versus the IReST reading speed. A linear regression analysis indicated that only maximum reading speed was accepted as a valid predictor for the IReST reading speed under the corrected nearvision condition when a stepwise method was used (R = 0.85, adjusted $R^2 = 0.72$, P < .001). Entering critical print size and reading acuity did not cause a significant improvement in R^2 .

TABLE 1.

C a rin A as $(\pm D)$ at sale rin in rin C	- EAD
---	-------

		Cha A	Cha B	Cha C	O a (95% C f , B a)
aga (gA)		0.16 ± 0.04	0.17 ± 0.06	0.16 ± 0.04	0.16 ± 0.05 (0.02)
Ca (gA)		0.23 ± 0.06	0.25 ± 0.06	0.24 ± 0.06	$0.24 \pm 0.06 \ (0.02)$
a ag g (haa /	/)	270.87 ± 35.44	271.41 ± 34.49	278.03 ± 33.84	273.44 ± 34.37 (14.40)



Ba∙_A h а ga ha (ha A .ha B, ha А ha C,a 🏻 ha B . ha C). h а / ha h 95% lag h 1 h а а he а h 95% 1 28

Equation 1 is the regression model for the corrected near-vision condition.

IReST reading speed
=
$$1.21$$
(maximum reading speed) - 25.53 [1]

Under the uncorrected near-vision condition, both critical print size and maximum reading speed contributed to the

variance in the IReST reading speed (R = 0.84, adjusted $R^2 = 0.68$, P < .001). Equation 2 is the regression model:

IReST reading speed = 0.56(maximum reading spead) -97.04(critical prise size) + 199.80 [2]

The uncorrected near acuity was significantly correlated with

reading acuity and critical print size of the C-READ (r = 0.89,

TABLE 2.

h C	ſ	D# a	ia a⊾ _i h,i1	Ba	h a a a	C-EAD a a	1
a a	aa	0		Cha	O v v ×	Cha Cha O	E., E.
a, ga		90	6.5% 0%	0%	0%	3.5%	0%
Ca		6.	5.1% 0%	0%	32.5%	2.4%	0%
а	a g	▲ 8 ⁻	1.7% 7.5%	0%	10.1%	0.7%	0%

Optometry and Vision Science, Vol. 94, No. 6, June 2017

Copyright © American Academy of Optometry. Unauthorized reproduction of this article is prohibited.

P < .001, for reading acuity, and r = 0.91, P < .001, for critical print size) and with the IReST reading speed (r = -0.63, P < .001), but was insignificantly correlated with maximum reading speed (r = -0.34, P = .06).

DISCUSSION

Written Chinese is quite different from alphabetic languages in several important ways. The final design of the C-READ was the outcome of increasingly deeper understanding of the impacts of these unique characteristics of written Chinese on reading. In particular, not only the sentence-level spatial complexities, quantifully corrected for the reading distance, there was only a small variation in their critical print sizes. Moreover, the font size of the text passage, 1.1 logMAR, exceeded the critical print sizes by at least 0.66 log unit. With this substantial amount of acuity reserve, the observers' passage reading speeds could be predicted solely by their C-READ maximum reading speed (Equation 1). In comparison, when the refractive errors were not corrected, the variance accounted for by critical print size was large. The larger the critical print size, the smaller the acuity reserve for reading and the slower the reading speed. In fact, the critical print size was only 0.1 to 0.2 log unit larger than 1.1 logMAR in six observers, 0.0 to 0.1 log unit larger in four observers, equal to 1.1 logMAR in two observers, and larger than 1.1 logMAR in two observers. Therefore, a large portion of the observers had small or no acuity reserves when they read the text passages with uncorrected vision. Consequently, critical print size, a parameter related to the acuity reserve, became the major predictor of the text passage reading speed (Equation 2).

The C-READ has good interchart reliability. Ninety-five percent of the repeated measurement errors would fall within 0.04 logMAR, 0.12 logMAR, and 19.2 characters per minute from the true reading acuity, critical print size, and maximum reading speed, respectively.²⁶ Because the interchart error for the C-READ is only a few characters for reading acuity, is approximately one line for critical print size, and is no larger than 7% of the average maximum reading speed, clinicians should feel

۳B

C-EAD confident to use the three charts interchangeably. They can use different charts to test the left eye, right eye, and both eyes of one patient or to monitor the change of reading performance at up to three time points.

The variance component analysis of the C-READ parameters showed that interobserver variability accounted for most variance in reading acuity and maximum reading speed (Table 2), indicating that the charts can detect individual differences in reading acuity and speed. This is similar to the finding of Stifter and colleagues,²⁹ who validated the Radner Reading charts. However, in contrast to Stifter and colleagues'29 study, which found that the majority of the variance in critical print size came from "unidentified sources," two-thirds of the C-READ critical print size variance came from the observers, and one-third came from the interaction between the observers and the three versions of the C-READ. Because critical print size plays an important role in determining the magnification of the reading aid for a patient with visual impairment,¹⁴ clinicians may consider testing reading twice using two charts to improve the accuracy of critical print size assessment.

Within-subjects comparison of MNRead-English and C-READ is difficult, because of the obvious language barrier in performing reading tasks using one's native tongue and a foreign language. However, between-subjects comparisons of native English readers reading MNRead-English and native Chinese readers reading C-READ may help illustrate the cross-language



Optometry and Vision Science, Vol. 94, No. 6, June 2017

differences in reading assessment. In a recent study, Calabrèse et al.³⁰ compiled a large set of MNRead–English data from English readers of a wide range of ages. From their age regression models, it was estimated that MNRead reading acuity, critical print size, and maximum reading speed for age 22 were $-0.168 \log MAR$, 0.077 logMAR, and 201.6 words per minute, respectively. In comparison, the reading acuity and critical print size for the Chinese readers, as determined by the C-READ, were 0.34 and 0.16 logMAR larger than the corresponding MNRead parameters, respectively (Table 1). This was not surprising because we have shown that visual acuity obtained using Roman letters (Sloan letters) was one line better than that obtained using two- to four-stroke simple Chinese characters and that the difference became even larger with more complex Chinese characters.²⁰ The difference between the MNRead-English maximum reading speed (202 words per minute) and that of the C-READ (273 characters per minute) was more difficult to comprehend because a Chinese word may contain one or more characters. On the other hand, the MNRead sentences have 10 six-character "standard-length" words with 12 to 15 syllables. The maximum reading speed of 202 words per minute of the MNRead can thus be converted to 242 to 303 syllables per minute. Because one Chinese character is one syllable, the reading speed of 273 characters per minute of the C-READ is 273 syllables per minute. Therefore, MNRead-English and C-READ maximum reading speed are comparable if the number of syllables uttered per minute is considered.

The C-READ is made of simplified Chinese characters, which are read by the majority of Chinese readers. Recently, Cheung et al.³¹ and Cheong et al.³² developed a logarithmic reading acuity chart for traditional Chinese readers, who reside mainly in Taiwan and Hong Kong. This reading chart differs from the C-READ in several important ways. First, many simplified Chinese characters have much fewer strokes and thus are visually simpler and have lower spatial frequency components than their traditional counterparts. The following are one of the C-READ sentences (89 strokes) and its traditional rendering (121 strokes):

我们都来画一画家乡的景物 (simplified Chinese)

我們都來畫壹畫家鄉的景物 (traditional Chinese)

Notice the difference between characters "一" and "壹" (one) and "乡" and "鄉" (home land). The difference in spatial complexity between simplified and traditional Chinese is much greater than that among most languages using Roman alphabets. As mentioned in Methods, special considerations had to be given to the presence of very simple characters in the C-READ sentences to ensure within- and between-charts consistency. Because characters with more strokes are known to have larger recognition threshold sizes,²⁰ the low vision aid magnification or school textbook font size determined using the C-READ critical print size is likely to be too small for patients or school pupils who read traditional Chinese. Indeed, the C-READ critical print size for 22-year-olds (Table 1) was smaller than that measured using the traditional Chinese charts $(0.24 \pm 0.06 \log MAR \text{ vs. } 0.51 \log MAR)$.³² Finally, many Chinese do not read simplified and traditional Chinese with equal fluency. It depends on where they are brought up. In summary, the simplified and traditional Chinese reading charts should be considered as two different instruments for two different populations.

This study examined only the effects of refractive errors on the C-READ in a group of highly uniform young normal observers.

Other pathological conditions, such as cataract, macular degeneration, glaucoma, and hemianopia, are known to affect reading performance but for different reasons. A recent study of reading performance in normally sighted English readers of different ages demonstrated significant age dependence of the MNRead–determined maximum reading speed, critical print size, and reading acuity.³⁰ Future studies are needed to expand the scope of the applications of the C-READ to these conditions.

ACKNOWLEDGMENTS

The authors thank Dr. Gordon Legge for his invaluable advices. Received July 22, 2016; accepted April 4, 2017.

REFERENCES

- Ahn SJ, Legge GE. Psychophysics of reading—XIII. Predictors of magnifier-aided reading speed in low vision. Vision Res 1995;35:1931–8.
- Legge GE, Ross JA, Isenberg LM, et al. Psychophysics of reading. Clinical predictors of low-vision reading speed. Invest Ophthalmol Vis Sci 1992;33:677–87.
- Runge PE. Eduard Jaeger's Test-Types (Schrift-Scalen) and the historical development of vision tests. Trans Am Ophthalmol Soc 2000;98:375–438.
- Sloan LL, Brown DJ. Reading cards for selection of optical aids for the partially sighted. Ame J Ophthalmol 1963;55:1187–99.
- Mansfield JS, Ahn SJ, Legge GE, et al. A new reading-acuity chart for normal and low vision. In: Ophthalmic and Visual Optics/ Noninvasive Assessment of the Visual System Technical Digest, vol. 3. Washington, DC: Optical Society of America; 1993;232–5.
- Radner W, Willinger U, Obermayer W, et al. A new reading chart for simultaneous determination of reading vision and reading speed. Klin Monbl Augenheilkd 1998;213:174–81.
- Colenbrander A. Measuring vision and vision loss. In: Duane TD, Tasman W, Jaeger EA, eds. Duane's Clinical Ophthalmology [on CD-ROM], Philadelphia, PA: Lippincott Williams & Wilkins; 2001.
- 8. Radner W, Obermayer W, Richter-Mueksch S, et al. The validity and reliability of short German sentences for measuring reading speed. Graefes Arch Clin Exp Ophthalmol 2002;240:461–7.
- 9. Idil SA, Caliskan D, Idil NB. Development and validation of the Turkish version of the MNREAD visual acuity charts. Turk J Med Sci 2011;41:565–70.
- Maaijwee KJ, Meulendijks CF, Radner W, et al. The Dutch version of the Radner Reading Chart for assessing vision function. Ned Tijdschr Geneeskd 2007;151:2494–7.
- Mataftsi A, Bourtoulamaiou A, Haidich AB, et al. Development and validation of the Greek version of the MNREAD acuity chart. Clin Exp Optom 2013;96:25–31.
- Alió JL, Radner W, Plaza-Puche AB, et al. Design of short Spanish sentences for measuring reading performance: Radner-Vissum test. J Cataract Refract Surg 2008;34:638–42.
- Castro CT, Kallie CS, Salomão SR. Development and validation of the MNREAD reading acuity chart in Portuguese. Arq Bras Oftalmol 2005;68:777–83.
- 14. Legge GE. Psychophysics of Reading in Normal and Low Vision, Mahwah, NJ: Lawrence Erlbaum Associates Publishers; 2007.
- 15. Ishii M, Seki M, Harigai R, et al. Reading performance in patients with glaucoma evaluated using the MNREAD charts. Jpn J Ophthalmol 2013;57:471–4.

- 724 Ch 🚓 g A Ha a.
- Hahn GA, Penka D, Gehrlich C, et al. New standardised texts for assessing reading performance in four European languages. Br J Ophthalmol 2006;90:480–4.
- Trauzettel-Klosinski S, Dietz K, IReST Study Group. Standardized assessment of reading performance: the new International Reading Speed Texts IReST. Invest Ophthalmol Vis Sci 2012;53:5452–61.
- Neubig G, Duh K. How much is said in a tweet? A multilingual, information-theoretic perspective. In: Hovy E, Markman V, Martell C, et al, eds. AAAI Spring Symposium on Analyzing Microtext; 2013, Stanford, CA: The AAAI Press; 2013;32–9 Available at: http:// www.aaai.org/ocs/index.php/SSS/SSS13/paper/viewFile/5698/ 5906. Accessed April 4, 2017.
- Liao HT, Fu KW, Hale SA. How much is said in a microblog? A multilingual inquiry based on Weibo and Twitter. In: de Roure D, Burnap P, Halford S, eds. ACM Web Science Conference, WebSci 2015, June 28–July 1, 2015. Oxford, UK: ACM; 2015:1–25:9. Available at: http://arxiv.org/abs/1506.00572. Accessed April 4, 2017.
- Zhang JY, Zhang T, Xue F, et al. Legibility variations of Chinese characters and implications for visual acuity measurement in Chinese reading population. Invest Ophthalmol Vis Sci 2007;48:2383–90.
- 21. [Compulsory Education Curriculum Standard Experimental Textbook]. Beijing, China: 人民教育出版社 [People's Education Press]; 2001.
- 22. Maaijwee K, Mulder P, Radner W, et al. Reliability testing of the Dutch version of the Radner Reading Charts. Optom Vis Sci 2008; 85:353–8.
- 23. Shu H, Chen X, Anderson RC, et al. Properties of school Chinese: implications for learning to read. Child Dev 2003;74:27–47.
- Cheung SH, Kallie CS, Legge GE, et al. Nonlinear mixed-effects modeling of MNREAD data. Invest Ophthalmol Vis Sci 2008; 49:828–35.

- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. Lancet 1986;1: 307–10.
- 26. Bland JM, Altman DG. Measurement error. BMJ 1996;313:744.
- Cicchetti DV. Guidelines, criteria, and rules of thumb for evaluating normed and standardized assessment instruments in psychology. Psychol Assess 1994;6:284–90.
- Liu L, Klein SA, Xue F, et al. Using geometric moments to explain human letter recognition near the acuity limit. J Vis 2009;9:26.1–18.
- 29. Stifter E, König F, Lang T, et al. Reliability of a standardized reading chart system: variance component analysis, test-retest and inter-chart reliability. Graefes Arch Clin Exp Ophthalmol 2004;242:31–9.
- Calabrèse A, Cheong AM, Cheung SH, et al. Baseline MNREAD measures for normally sighted subjects from childhood to old age. Invest Ophthalmol Vis Sci 2016;57:3836–43.
- Cheung JP, Liu DS, Lam CC, et al. Development and validation of a new Chinese reading chart for children. Ophthalmic Physiol Opt 2015;35:514–21.
- Cheong AM, Legge GE, Liu L. Validation of a new Chinese reading acuity chart for clinical research. Optom Vis Sci 2012;89 E-abstract 125172.

Lei Liu

School of Optometry University of Alabama at Birmingham 1716 University Blvd Birmingham, AL 35294 e-mail: liul7788@uab.edu; or Cong Yu

School of Psychological and Cognitive Sciences Peking University Beijing 100181, China e-mail: yucong@pku.edu.cn